POWER SENSORS Precision Filtered RF

4027F SERIES



The Bird 4027F Series RF Power Sensors are designed specifically for use in semiconductor processing and other precision process applications where the effects of amplitude modulation and harmonics need to be eliminated from the measurement. Because accuracy is critical, these sensors are capable of $\pm 1\%$ accuracy at the calibrated frequency and power level. Calibration is traceable to the National Institute of Standards and Technology (NIST), providing additional confidence in sensor measurement.

±1% Accuracy, Precision
Directional RF Power Sensors

PRODUCT FEATURES

- Provides filtered measurement where the effects of amplitude modulation and harmonics need to be eliminated
- ±1% accuracy at specified frequencies and power levels minimizing poor production yields
- Models do not need to be field calibrated before use and only need to be calibrated once every six months
- Plug and Play with the 4421A Series Multifunction Power Meter
- Dozens of connector options available
- Calibration traceable to the National Institute of Standards and Technology

FOR USE WITH

- 4421A Multifunction Power Meters
- SCC7 Calibration Cart
- MSCC7 Calibration Cart



PRECISION FILTERED RF POWER SENSORS

4027F SERIES

Specifications

PRODUCT SELECTION GUIDE

| Model Number | Freq Range | Power Range | Calibration Frequency, Typ | Harmonic Rejection, Min | LF Rejection | Max Error Induced by 10% AM |
|--------------|----------------|----------------|--------------------------------------|---|-----------------|----------------------------------|
| 4027F2M | 1.8 to 2.2 MHz | 100 W to 10 kW | 1.8, 2.0, 2.17 MHz | 26 dB@3.6 MHz to 3.8 MHz, 30 dB@>3.8 MHz | Not specified | 0.2%@<5 kW 1%@5kW to 10 kW |
| 4027F10M | 12 to 15 MHz | 100W to 10kW | 12.0, 12.5, 13.56, 14.0, 15.0 MHz | 30 dB@>25 MHz | 30 dB@<1 MHz | 0.2%@<5 kW 1%@5kW to 10 kW |
| 4027F60M | 57 to 63 MHz | 100W to 6 kW | 57.0, 58.5, 60.0, 61.5, 63 MHz | 30 dB@>114.0 MHz | 30 dB@<15.0 MHz | 0.2%@<1.5 kW 1%@1.5kW to 3 kW |

MEASUREMENT

| Accuracy | \pm 1% at calibration frequencies and power levels \pm 2% at other frequency and power levels Add 2% to uncertainty outside 25 \pm 5 °C |
|-------------------------|---|
| Uniformity | 2% maximum unit to unit, at calibration frequency and power levels |
| Speed | 2 readings per second |
| VSWR Range | 1.00 to 2.00 |
| Directivity | 28 dB |
| Insertion Loss | <0.05 dB |
| Calibration Power Level | 1700 W |
| Maximum Power | 10 kW units: 12 kW max 6 kW units : 7.2 kW max |

CONNECTORS

RF Connectors Customer specified

SYSTEM

| Recommended Calibration Interval | 6 month. Performance before and performance after data to be supplied for units |
|-------------------------------------|---|
| Power Supply | External DC, 12 VDC, supplied from Bird 4421 Power Meter |

ENVIRONMENTAL

| Operating Temperature | 0 °C to 50 °C (32 °F to 122 °F) |
|-----------------------|-----------------------------------|
| Storage Temperature | -20 °C to 70 °C (-4 °F to 158 °F) |
| Humidity | 95% max (non-condensing) |
| Altitude | Up to 10,000 ft (3,048 m) |

PHYSICAL

| Size | 5.2 in x 2.5 in x 3.25 in (137 mm x 64 mm x 83 mm) |
|--------|---|
| Weight | 1 lb 13 oz (0.8 kg) |

CERTIFICATIONS

| EMC | Designed to carry CE mark | |
|-----|---------------------------|--|
| | | |

birdrf.com/products

The **RF** Experts | USA Sales : 30303 Aurora Rd, Solon, OH 44139 | www.birdrf.com Phone: +1 440.248.1200 / 866.695.4569 [Toll Free]

ORDERING GUIDE

Example: 4027F10M-0102

4027F10M with a N(f) Input Connector and a N(m) Output Connector

| Sensor | - | Input Connector | Output Connector |
|--------|---|-----------------|------------------|
|--------|---|-----------------|------------------|

| <u> </u> | ector |
|---------------------------------|-------|
| 7F2M 01 = 4240-062, N(f) | |
| 7F10M 02= 4240-063, N(m) | |
| 7F60M 04= 4240-100, C(f) | |
| 11= 4240-096, (1 5/8 fix | xed) |
| 12= 4240-268, HN(f) | |
| 13= 4240-278, HN(f) | |
| 14= 4240-344, 7/16(f) | |
| 15 = 4240-363, 7/16(m) | |
| 16= 4240-370, SQS(m) | |
| 17= 4240-371, SQS(f) | |
| 18= 4240-372, SQS(m- | p) |
| 19 = 4240-376, QRM(f) | |
| 20 = 4240-374, QDS-UL | (m) |
| 21 = 4240-375, SQS(f-p) |) |
| 22 = 4240-373, QDS-UL | (f) |
| 23 = 4240-377, QRM(m) |) |
| 24 = 4240-377-2, QRM(| f-p) |
| 25 = 4240-378, QRM(m- | -p) |
| 26 = 4240-376-20, GQL | (f) |
| 27 = 4240-377-20, GQL | (m) |
| 28 = 4240-376-10, GQN | 1(f) |
| 29 = 4240-377-10, GQN | 1(m) |
| 30 = 4240-208, 1-5/8 Sv | vivel |
| 34 = 4240-031, LC(f) | |
| 35 = 4240-025, LC(m) | |
| 36 = 4240-002, 7/8 | |
| 39 = 4240-373-2, QDS(f |) |
| 40 = 4240-374-2, QDS(r | n |

